









































Product code	Appearance	Description	Units per package
BT10		Self-contained Biological Indicator for EO sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per unit.	100
BT10		Self-contained Biological Indicator for EO sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per unit.	30
BT20		Self-contained Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per unit.	100
BT20		Self-contained Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per unit.	30
BT30		Semi Self-contained Biological Indicator for Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per vial. Each box of 50 tubes allows 25 detections: 1 spore tube + 1 culture medium tube.	50
BT80		Self-contained Biological Indicator for Moist Heat decontamination process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁴ spores per unit.	100
BT100		Self-contained Biological Indicator for Formaldehyde sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per vial.	100
BT91		Self-contained Biological Indicator for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per vial.	100
BT90 F1		Rapid Readout Colorimetric System: Self-contained Biological Indicator for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per unit. Readout: 8 hours.	100
BT200 F1		Rapid Readout Colorimetric System: Self-contained Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per unit. Readout: 8 hours.	100
BT101 F1		Rapid Readout Colorimetric System: Self-contained Biological Indicator for Formaldehyde sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per unit. Readout: 8 hours.	100
BT220		Rapid Readout System: Self-contained Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ spores per unit. Readout: 3 hours.	50
BT221		Rapid Readout System: Self-contained Biological Indicator for Monitoring of gravity displacement steam sterilization cycles at 132 °C (270 °F). <i>Geobacillus stearothermophilus</i> (ATCC 9372) 10 ⁵ or 10 ⁶ spores per unit. Readout: 1 hour.	50
BT110		Rapid Readout System: Self-contained Biological Indicator for EO sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per unit. Readout: 4 hours.	50




Product code	Appearance	Description	Units per package
BT40/3		Spore Strips Biological Indicator for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ³ spores per strip.	100
BT40/4		Spore Strips Biological Indicator for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁴ spores per strip.	100
BT40/5		Spore Strips Biological Indicator for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁵ spores per strip.	100
BT40/6		Spore Strips Biological Indicator for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per strip.	100
BT40/7		Spore Strips Biological Indicator for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁷ spores per strip.	100
BT50/3		Spore Strips Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ³ spores per strip.	100
BT50/4		Spore Strips Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁴ spores per strip.	100
BT50/5		Spore Strips Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per strip.	100
BT50/6		Spore Strips Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per strip.	100
BT50/7		Spore Strips Biological Indicator for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁷ spores per strip.	100
BT60		Dual Spore Strips Biological Indicator for Steam, Ethylene Oxide and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ or 10 ⁶ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per strip.	100
BT70/5		Spore Strips Biological Indicator for Gamma Radiation sterilization process. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁵ spores per strip.	100
BT70/6		Spore Strips Biological Indicator for Gamma Radiation sterilization process. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁶ spores per strip.	100
BT70/7		Spore Strips Biological Indicator for Gamma Radiation sterilization process. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁷ spores per strip.	100




Product code	Appearance	Description	Units per package
BT92/5		Spore Strips Biological Indicator for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per strip.	100
BT92/6		Spore Strips Biological Indicator for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per strip.	100













Mini spore strips and spore discs biological indicators BIONOVA®

Product code	Appearance	Description	Units per package
BTD40-3		Spore Discs Biological Indicators for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/3 inches (8,46 mm)	100
BTD40-4		Spore Discs Biological Indicators for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/4 inches (6,35 mm)	100
BTD40-8		Spore Discs Biological Indicators for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/8 inches (3,18 mm)	100
BTD50-3/5		Spore Discs Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per disc. Diameter: 1/3 inches (8,46 mm)	100
BTD50-3/6		Spore Discs Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per disc. Diameter: 1/3 inches (8,46 mm)	100
BTD50-4/5		Spore Discs Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per disc. Diameter: 1/4 inches (6,35 mm)	100
BTD50-4/6		Spore Discs Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per disc. Diameter: 1/4 inches (6,35 mm)	100
BTD50-8/5		Spore Discs Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per disc. Diameter: 1/8 inches (3,18 mm)	100
BTD50-8/6		Spore Discs Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per disc. Diameter: 1/8 inches (3,18 mm)	100
BTD60-3/5		Dual Spore Discs Biological Indicators for EO, Steam and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/3 inches (8,46 mm)	100
BTD60-3/6		Dual Spore Discs Biological Indicators for EO, Steam and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/3 inches (8,46mm)	100
BTD60-4/5		Dual Spore Discs Biological Indicators for EO, Steam and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/4 inches (6,35 mm)	100
BTD60-4/6		Dual Spore Discs Biological Indicators for EO, Steam and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/4 inches (6,35 mm)	100
BTD60-8/5		Dual Spore Discs Biological Indicators for EO, Steam and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/8 inches (3,18 mm)	100


Product code	Appearance	Description	Units per package
BTD60-8/6		Dual Spore Discs Biological Indicators for EO, Steam and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per disc. Diameter: 1/8 inches (3,18 mm)	100
BTD90-3/6		Spore Discs Biological Indicators for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per disc. Diameter: 1/3 inches (8,46 mm)	100
BTD90-4/6		Spore Discs Biological Indicators for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per disc. Diameter: 1/4 inches (6,35 mm)	100
BTD90-8/6		Spore Discs Biological Indicators for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per disc. Diameter: 1/8 inches (3,18 mm)	100
BTM40.2-10		Mini Spore Strips Biological Indicators for EO and Dry Heat sterilization process. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per strip (2 mm x 10 mm)	100
BTM40G/6		Mini Spore Strips Biological Indicators for EO and Dry Heat sterilization process, in single glassine packaging. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per strip (2 mm x 10 mm)	100
BTM50.2-10/5		Mini Spore Strips Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per strip (2 mm x 10 mm)	100
BTM50.2-10/6		Mini Spore Strips Biological Indicators for Steam sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per strip (2 mm x 10 mm)	100
BTM60.2-10/5		Dual Mini Spore Strips Biological Indicators for Steam, Ethylene Oxide and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per strip (2 mm x 10 mm)	100
BTM60.2-10/6		Dual Mini Spore Strips Biological Indicators for Steam, Ethylene Oxide and Dry Heat sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ and <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores per strip (2 mm x 10 mm).	100




Product code	Appearance	Description	Units per package
BTC90-1/4		Perforated Stainless Steel Coupons Biological Indicators for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁴ spores per coupon.	100
BTC90-1/5		Perforated Stainless Steel Coupons Biological Indicators for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per coupon.	100
BTC90-1/6		Perforated Stainless Steel Coupons Biological Indicators for VH ₂ O ₂ sterilization process. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per coupon.	100













Product code	Appearance	Description	Units per package
IC10/20		Dual Incubator for Biological Indicators 37±2 °C or 60±2 °C.	1
IC10/20F		Auto-reader for Rapid Readout Self-Contained Biological Indicators 37±2 °C or 60±2 °C.	1
TB-IC10/20		Temperature sensing probe. Accessory to IC10/20.	1

Product code	Appearance	Description	Units per package
MC1030-1		Culture Medium with pH indicator for the <i>Bacillus atrophaeus</i> (ATCC 9372) growth (1 ml). Turns from blue to yellow.	50
MC1030-80		Culture Medium with pH indicator for the <i>Bacillus atrophaeus</i> (ATCC 9372) growth (80 ml). Turns from blue to yellow.	1
MC1030-150		Culture Medium with pH indicator for the <i>Bacillus atrophaeus</i> (ATCC 9372) growth (150 ml). Turns from blue to yellow.	1
MC1030-500		Culture Medium with pH indicator for the <i>Bacillus atrophaeus</i> (ATCC 9372) growth (500 ml). Turns from blue to yellow.	1
MC20-1		Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) growth (1 ml). Turns from purple to yellow.	50
MC20-80		Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) growth (80 ml). Turns from purple to yellow.	1
MC20-150		Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) growth (150 ml). Turns from purple to yellow.	1
MC20-500		Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) growth (500 ml). Turns from purple to yellow.	1
MC1020-1		Dual Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) and <i>Bacillus atrophaeus</i> (ATCC 9372) growth (1 ml). Turns from red to yellow.	50
MC1020-80		Dual Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) and <i>Bacillus atrophaeus</i> (ATCC 9372) growth (80 ml). Turns from red to yellow.	1
MC1020-150		Dual Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) and <i>Bacillus atrophaeus</i> (ATCC 9372) growth (150 ml). Turns from red to yellow.	1
MC1020-500		Dual Culture Medium with pH indicator for the <i>Geobacillus stearothermophilus</i> (ATCC 7953) and <i>Bacillus atrophaeus</i> (ATCC 9372) growth (500 ml). Turns from red to yellow.	1
MC70-1		Culture Medium with pH indicator for the <i>Bacillus pumilus</i> (ATCC 27142) growth (1 ml). Turns from red to yellow.	50

Product code	Appearance	Description	Units per package
BT10S/2		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ¹ spores/0,1 ml.	1 (10ml)
BT10S/3		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ² spores/0,1 ml.	1 (10ml)
BT10S/4		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ³ spores/0,1 ml.	1 (10ml)
BT10S/5		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁴ spores/0,1 ml.	1 (10ml)
BT10S/6		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁵ spores/0,1 ml.	1 (10ml)
BT10S/7		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁶ spores/0,1 ml.	1 (10ml)
BT10S/8		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁷ spores/0,1 ml.	1 (10ml)
BT10S/9		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Ethylene Oxide sterilization. <i>Bacillus atrophaeus</i> (ATCC 9372) 10 ⁸ spores/0,1 ml.	1 (10ml)
BT20S/2		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ¹ spores/0,1 ml.	1 (10ml)
BT20S/3		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ² spores/0,1 ml.	1 (10ml)
BT20S/4		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ³ spores/0,1 ml.	1 (10ml)
BT20S/5		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁴ spores/0,1 ml.	1 (10ml)
BT20S/6		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores/0,1 ml.	1 (10ml)
BT20S/7		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores/0,1 ml.	1 (10ml)

Product code	Appearance	Description	Units per package
BT20S/8		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁷ spores/0,1 ml.	1 (10ml)
BT24S/2		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ¹ spores /0,1 ml.	1 (10ml)
BT24S/3		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ² spores /0,1 ml.	1 (10ml)
BT24S/4		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ³ spores /0,1 ml.	1 (10ml)
BT24S/5		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁴ spores /0,1 ml.	1 (10ml)
BT24S/6		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁵ spores /0,1 ml.	1 (10ml)
BT24S/7		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁶ spores /0,1 ml.	1 (10ml)
BT24S/8		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁷ spores /0,1 ml.	1 (10ml)
BT24S/9		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Steam sterilization. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁸ spores /0,1 ml.	1 (10ml)
BT70S/2		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ¹ spores/0,1 ml.	1 (10ml)
BT70S/3		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ² spores/0,1 ml.	1 (10ml)
BT70S/4		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ³ spores/0,1 ml.	1 (10ml)
BT70S/5		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁴ spores/0,1 ml.	1 (10ml)
BT70S/6		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁵ spores/0,1 ml.	1 (10ml)

Product code	Appearance	Description	Units per package
BT70S/7		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁶ spores/0,1 ml.	1 (10ml)
BT70S/8		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁷ spores/0,1 ml.	1 (10ml)
BT70S/9		Spore Suspension of bacterial spores for direct inoculation of products or the preparation of customized biological indicators for Gamma Radiation. <i>Bacillus pumilus</i> (ATCC 27142) 10 ⁸ spores/0,1 ml.	1 (10ml)

Product code	Appearance	Description	Units per package
BT21/4		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁴ spores per ampoules (7 mm x 40 mm)	200 (+) 20 (-)
BT21/5		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per ampoules (7 mm x 40 mm)	200 (+) 20 (-)
BT21/6		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per ampoules (7 mm x 40 mm)	200 (+) 20 (-)
BT22/4		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁴ spores per ampoules (7 mm x 25 mm)	250 (+) 25 (-)
BT22/5		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per ampoules (7 mm x 25 mm)	250 (+) 25 (-)
BT22/6		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per ampoules (7 mm x 25 mm)	250 (+) 25 (-)
BT23/4		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁴ spores per ampoules. (12 mm x 55 mm). 1 ml of culture medium.	50 (+) 5 (-)
BT23/5		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁵ spores per ampoules. (12 mm x 55 mm). 1 ml of culture medium.	50 (+) 5 (-)
BT23/6		Self-contained Spores Ampoules for monitoring industrial Steam sterilization of liquids. <i>Geobacillus stearothermophilus</i> (ATCC 7953) 10 ⁶ spores per ampoules. (12 mm x 55 mm). 1 ml of culture medium.	50 (+) 5 (-)
BT24/4		Self-contained Spores Ampoules for monitoring industrial low temperature 110 - 118 °C Steam sterilization of liquids. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁴ spores per ampoules. (7 mm x 40 mm)	200 (+) 20 (-)
BT24/5		Self-contained Spores Ampoules for monitoring industrial low temperature 110 - 118 °C Steam sterilization of liquids. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁵ spores per ampoules. (7 mm x 40 mm)	200 (+) 20 (-)
BT24/6		Self-contained Spores Ampoules for monitoring industrial low temperature 110 - 118 °C Steam sterilization of liquids. <i>Bacillus subtilis</i> DSM 5230 (ATCC 35021), 10 ⁶ spores per ampoules. (7 mm x 40 mm)	200 (+) 20 (-)